










GIRFReco.jl: An Open-Source Pipeline for Spiral Magnetic Resonance Image (MRI) Reconstruction in Julia

Alexander Jaffray ^{1,2*} ^{1*} ³, Kâmil Uludağ ^{1,4,5}, and Lars Kasper ¹

1 Krembil Research Institute, University Health Network, Ontario, Canada **2** MRI Research Centre, University of British Columbia, Vancouver, Canada **3** Department of Physics, Norwegian University of Science and Technology, Trondheim, Norway **4** Department of Medical Biophysics, University of Toronto, Canada **5** Physical Sciences, Sunnybrook Research Institute, Sunnybrook Health Sciences Centre, Toronto, ON, Canada  Corresponding author * These authors contributed equally.

DOI: [10.21105/joss.05877](https://doi.org/10.21105/joss.05877)

Software

- [Review](#) 
- [Repository](#) 
- [Archive](#) 

Editor: [Kevin M. Moerman](#)  

Reviewers:

- [@cncastillo](#)
- [@mathieboudreau](#)
- [@felixhoger](#)

Submitted: 01 June 2023

Published: 02 May 2024

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](https://creativecommons.org/licenses/by/4.0/)).

Summary

Magnetic Resonance Imaging (MRI) acquires data in the frequency domain (k-space), with the sampling pattern traversed by a path known as the k-space trajectory. It is desirable to implement MRI data sampling using k-space trajectories with high acquisition efficiency (i.e., a fast coverage of k-space). Traditional Cartesian MRI traverses k-space by acquiring individual lines of the k-space, each requiring an excitation, a phase-encoding step, and a short readout gradient. However, it is possible to traverse k-space with an arbitrary trajectory, achieved by a long sequence of readout gradients, thus presenting the opportunity to acquire more sampling points per excitation. Spiral trajectories are a popular and efficient method for traversing k-space with a long readout, as well as classical echo-planar imaging (EPI) trajectories. Non-Cartesian trajectories, such as spiral trajectories, yield significant reductions in the number of excitations required for the acquisition of an image, thus offering considerable acceleration and improvements in signal-to-noise ratio (SNR) per unit time at the cost of reconstruction complexity. These improvements are particularly beneficial in diffusion MRI due to sequence timing constraints ([Lee et al., 2021](#)).

The actual k-space trajectory applied during the MRI experiment can differ from the nominal trajectory due to hardware imperfections, resulting in image artifacts such as ghosting, blurring or geometric distortion. This problem is exacerbated in many non-Cartesian trajectories, such as spirals, because these fast imaging protocols place high demands on the gradient hardware of the MRI system ([Block & Frahm, 2005](#)). Accurate characterization of the system hardware is necessary and can be used for k-space trajectory correction, for example via a gradient impulse response function (GIRF) ([Addy et al., 2012](#); [Vannesjo et al., 2013](#)).

The high acquisition efficiency of non-Cartesian trajectories originates, in part, from the prolonged readout duration which allows for more samples to be acquired per excitation (single-shot or few-interleave scanning). However, when using such long readouts, the image encoding scheme is susceptible to static off-resonance (or field inhomogeneity, B_0), resulting in image artifacts that scale with readout duration. For non-Cartesian trajectories, these artifacts are difficult to correct in post-processing, but can be effectively addressed during image reconstruction by incorporating spatial off-resonance measurements into the signal model ([Sutton et al., 2003](#)).

Therefore, recent spiral imaging approaches often rely on an expanded signal model incorporating system imperfections and off-resonance maps ([Engel et al., 2018](#); [Graedel et al., 2021](#); [Kasper](#)

et al., 2018; Kasper et al., 2022; Lee et al., 2021; Robison et al., 2019; Vannesjo et al., 2016; Wilm et al., 2011, 2015), in combination with parallel imaging acceleration using multiple receiver coils and iterative non-Cartesian image reconstruction algorithms, e.g., CG-SENSE (Pruessmann et al., 2001).

Here, we introduce the open-source GIRFReco.jl reconstruction pipeline, which provides a single ecosystem implementation of this state-of-the-art approach to non-Cartesian MRI in the programming language Julia (Bezanson et al., 2017). The core reconstruction routines rely upon the public Julia package MRIReco.jl, a comprehensive open-source image reconstruction toolbox. To enable robust, accessible and fast MRI with spiral gradient waveforms, GIRFReco.jl is designed as an end-to-end signal processing pipeline, from open-standard raw MR data ([ISMR]MRD (Inati et al., 2017)) to final reconstructed images (NIFTI neuroimage data format (NIFTI, 2003)). It integrates system characterization information via GIRF correction for accurate representation of the encoding fields, relevant calibration data (coil sensitivity and static off-resonance maps) and iterative parallel imaging reconstruction for non-Cartesian k-space sampling patterns, including spiral trajectories.

Statement of Need

Existing open-source solutions for the correction of system imperfections and static off-resonance in MRI are often implemented within the framework of mature image reconstruction suites such as BART (Blumenthal et al., 2022), Gadgetron (Hansen & Sørensen, 2013) and MIRT (Fessler, n.d.).

However, the aforementioned complexity of the image reconstruction task for spiral MRI currently necessitates the integration of tools from multiple of these software suites in order to establish a performant and comprehensive image reconstruction workflow (e.g., (Veldmann et al., 2022)). With each tool being developed in different programming languages (C for BART; C++ for Gadgetron; MATLAB, C++ and C for MIRT, etc.), maintaining and extending such an image reconstruction pipeline then requires cross-language expertise, adding significant overhead and complexity to development. This presents a significant barrier to efficient and reproducible image reconstruction and limits software accessibility and sustainability, especially for users without software engineering backgrounds.

The programming language Julia (Bezanson et al., 2017) provides a practical solution to this multiple-language problem by using a high-level interface to low-level compiled code, i.e., enabling fast prototyping with limited resources in an academic setting, while delivering a near-industrial-level efficiency of code execution, all within a single development environment.

In this work, we introduce GIRFReco.jl (initial version presented at the annual meeting of ISMRM 2022 (Jaffray et al., 2022b)), which implements an end-to-end, self-contained processing and image reconstruction pipeline for spiral MR data completely in Julia. Based on the established MRIReco.jl package, GIRFReco.jl incorporates model-based corrections (Sutton et al., 2003; Vannesjo et al., 2016; Wilm et al., 2011, 2015) to achieve high-quality spiral MRI reconstructions. Specifically, this reconstruction pipeline combines several major steps: (1) ESPIRiT coil sensitivity map estimation (Uecker et al., 2014); (2) Robust off-resonance (B_0) map estimation (Funai et al., 2008; Lin & Fessler, 2020); (3) Computation of the applied non-Cartesian k-space trajectory using GIRF correction (Vannesjo et al., 2013, 2016); (4) Iterative non-Cartesian MRI reconstruction (CG-SENSE) with off-resonance correction (Knopp et al., 2009; Pruessmann et al., 2001). Considering software reusability and sustainability, (1) and (4) of the abovementioned steps are handled by MRIReco.jl, a comprehensive modular open-source image reconstruction toolbox in Julia. Step (2), the B_0 map estimation, was developed as a Julia package MRIFieldmaps.jl by the original authors (Lin & Fessler, 2020) with our contribution of implementing an alternative algorithm (Funai et al., 2008) in Julia. Finally, we implemented step (3), the GIRF correction, in an original Julia package MRIGradients.jl (Jaffray et al., 2022b), porting and refactoring the MATLAB code of the original authors

(Vannesjo & Graedel, 2020).

Functionality

Required Inputs

GIRFReco.jl requires raw MRI (k-space) data (in [ISMR]MRD format (Inati et al., 2017)) of the following scans as input:

1. Multi-echo Gradient-echo spin-warp (Cartesian) scan
 - must include at least two echo times (e.g., 4.92 ms and 7.38 ms at 3T)
2. Spiral scan
 - single or multi-interleave

At the moment, the slice geometry (thickness, field-of-view, and direction) of the Cartesian and spiral scans must be congruent, while the resolution does not need to be identical or isotropic.

Overview of Components

The following components are utilized within the spiral reconstruction pipeline of GIRFReco.jl (Fig. 1), and called from their respective packages. We indicate where the authors of GIRFReco.jl provided original contributions to the components by bold font.

1. Core iterative image reconstruction, using the Julia package MRIReco.jl
 - a. CG-SENSE (Pruessmann et al., 2001) algorithm for iterative non-Cartesian image reconstruction
 - b. ESPIRiT (Uecker et al., 2014) for sensitivity map estimation
2. Model-based correction
 - a. Static off-resonance (B_0 inhomogeneity) correction
 - **Smoothed B_0 map estimation, using an implementation of (Funai et al., 2008) and MRIFieldMaps.jl (Lin & Fessler, 2020)**
 - Static B_0 map correction, accelerated by a time-segmented implementation (Knopp et al., 2009) in MRIReco.jl (Knopp & Grosser, 2021)
 - b. Encoding field (trajectory) correction via Gradient impulse response function (GIRF) (Vannesjo et al., 2013)
 - Measurement with a phantom-based technique (Addy et al., 2012; Graedel et al., 2017; Robison et al., 2019)
 - **Estimation using open-source code (Wu et al., 2022)**
 - **Prediction via MRIGradients.jl (Jaffray et al., 2022b)**




Figure 1. Overview of the *GIRFReco.jl* signal processing and reconstruction pipeline. Depicted is the workflow from raw acquired k -space data to the final reconstructed images with the respective tasks (parallelogram), in/output data (rectangles) and the location of the processing components within different packages (colours).

Detailed Processing Pipeline

GIRFReco.jl executes the steps required (depicted in Figure 1) for spiral diffusion reconstruction in the following order:

1. Conversion of proprietary, vendor-specific raw image data to an open-source raw data format ([ISMR]MRD, (Inati et al., 2017)).
2. Reading of the trajectory or gradient sequence and synchronization of the k-space trajectory onto the time course of the sampled k-space data to resolve any sampling rate differences.
3. Model-based correction of the k-space sampling points (linear gradient self-terms) and data (k_0 eddy currents) using the gradient impulse response function (GIRF (Vannesjo et al., 2013), MRIGradients.jl (Jaffray et al., 2022b)).
4. Iterative reconstruction of Cartesian multi-echo gradient echo (GRE) scan.
5. Coil sensitivity map estimation (ESPIRiT (Uecker et al., 2014), MRIReco.jl) from the first echo of multi-echo Cartesian GRE data.
6. Off-resonance (B_0) map estimation and processing (MRIFieldmaps.jl, (Funai et al., 2008; Lin & Fessler, 2020)) based on multi-echo Cartesian GRE data.
7. Non-Cartesian, iterative parallel image reconstruction (cgSENSE) with off-resonance correction ((Knopp et al., 2009; Pruessmann et al., 2001), MRIReco.jl (Knopp & Grosser, 2021)).

Via dedicated configuration files, individual steps can be selectively applied or skipped during reconstruction, enabling assessment of the impact of different model-based corrections on final image quality. We demonstrate this use case by providing example reconstructions obtained from the GIRFReco.jl pipeline for a T_2 -weighted four-interleave spiral acquisition of a geometric structure phantom by the American College of Radiology (ACR). Reconstructions of both fully sampled and accelerated (using 1 of 4 interleaves, $R = 4$) datasets are depicted in Figure 2. *In vivo* brain images reconstructed from a T_2 -weighted single-interleave ($R=4$) spiral acquisition are presented in Figure 3 (Kasper et al., 2023). In all cases, improved image quality was obtained by successively increasing the complexity of the applied model-based corrections (nominal trajectory, added B_0 correction, GIRF-correction of gradients, GIRF correction of k_0 eddy currents). The improvements in quality are best seen when looking at high-contrast features of the images such as edges and corners, with subsequent corrections creating sharper edge contrast and reducing blurring of small features.

Note that the reconstruction results from the phantom experiments (both $R = 1$ and $R = 4$ reconstructions in Figure 2) can be fully reproduced using GIRFReco.jl and the corresponding dataset made publicly available (Jaffray et al., 2022a). For details, see the “Getting Started” section below.




Figure 2. Reconstructed four-interleave ($R=1$) and single-interleave ($R=4$) spiral images of a selected slice of the ACR phantom. Top row, from left to right: Images reconstructed from the nominal spiral gradient waveforms (“No Correction”), with correction for static off-resonance (“B₀ Correction”), B₀ + GIRF correction of the k-space trajectory (“B₀+GIRF Correction”), and additional correction for GIRF k₀ eddy currents (“B₀+GIRF+k₀ Correction”). Bottom row: Stepwise difference images between subsequent corrections.




Figure 3. Reconstructed in-vivo spiral images of a human brain images (single interleave, undersampling factor $R=4$). Top row: Images reconstructed from the nominal spiral gradient waveform (“No correction”), with correction for static off-resonance (“ B_0 map”), B_0 + GIRF correction of the k -space trajectory (“GIRF k_{xyz} ”), and additional correction for GIRF k_0 eddy currents (“Full Correction”). Bottom row: Consecutive absolute difference images of top-row reconstructions (5x scaled, i.e., $\pm 20\%$ max image intensity; or 50x scaled, i.e., $\pm 2\%$ max image intensity). A Cartesian image (echo 1 from the B_0 map scan) is used as the reference; its edges are overlaid to assess geometric congruency of the spiral images.

Quality of Life Features

In addition to providing an end-to-end reconstruction workflow, GIRFReco.jl offers methods for plotting images and calibration data at intermediate steps throughout the pipeline using PlotlyJS. Furthermore, intermediate reconstruction results, such as calculated coil sensitivity maps and B_0 maps are optionally stored as NIfTI files, a common neuroimaging data format supported by various analysis and visualization packages (NIFTI, 2003).

Getting Started

Up-to-date information about how to install GIRFReco.jl, run example reconstructions (e.g., reproducing Figure 2) and apply it to your own data can be found in the README.md provided in the GitHub repository. Further example scripts and technical documentation of GIRFReco.jl’s API, including its current feature set, is provided at ‘<https://brain-to.github.io/GIRFReco.jl>’, automatically generated by Documenter.jl.

Conclusion and Outlook

The presented pipeline, GIRFReco.jl, is an open-source end-to-end solution for spiral MRI reconstruction. It is developed in Julia, and allows users to obtain final images directly from raw MR data acquired by spiral k -space trajectories. Following best practices of software sustainability and accessibility, we rely on the established MR image reconstruction package MRIReco.jl in our pipeline, while extending its capability to handle the more complex use case of multiple model-based corrections, necessary for high-quality spiral MRI. Beyond spirals,

GIRFReco.jl can be readily utilized for data acquired under arbitrary non-Cartesian k-space trajectories; its features of model-based MRI reconstruction with GIRF and off-resonance corrections generalize to such sampling patterns in both 2D and 3D. Furthermore, GIRFReco.jl can be extended to handle additional model-based corrections (e.g., concomitant or higher-order encoding fields, (Bernstein et al., 1998; Vannesjo et al., 2016; Wilm et al., 2011, 2015)), and act as a self-contained template for generalized image reconstruction from raw scan and calibration data to interpretable and accessible images in Julia.

References

- Addy, N. O., Wu, H. H., & Nishimura, D. G. (2012). Simple method for MR gradient system characterization and k-space trajectory estimation. *Magnetic Resonance in Medicine*, 68(1), 120–129. <https://doi.org/10.1002/mrm.23217>
- Bernstein, M. A., Zhou, X. J., Polzin, J. A., King, K. F., Ganin, A., Pelc, N. J., & Glover, G. H. (1998). Concomitant gradient terms in phase contrast MR: Analysis and correction. *Magnetic Resonance in Medicine*, 39(2), 300–308. <https://doi.org/10.1002/mrm.1910390218>
- Bezanson, J., Edelman, A., Karpinski, S., & Shah, V. B. (2017). Julia: A Fresh Approach to Numerical Computing. *SIAM Review*, 59(1), 65–98. <https://doi.org/10.1137/141000671>
- Block, K. T., & Frahm, J. (2005). Spiral imaging: A critical appraisal. *Journal of Magnetic Resonance Imaging*, 21(6), 657–668. <https://doi.org/10.1002/jmri.20320>
- Blumenthal, M., Holme, C., Roeloffs, V., Rosenzweig, S., Schaten, P., Scholand, N., Tamir, J., Wang, X., & Uecker, M. (2022). *Mrirecon/bart: Version 0.8.00*. Zenodo. <https://doi.org/10.5281/ZENODO.592960>
- Engel, M., Kasper, L., Barmet, C., Schmid, T., Vionnet, L., Wilm, B., & Pruessmann, K. P. (2018). Single-shot spiral imaging at 7 T. *Magnetic Resonance in Medicine*, 80(5), 1836–1846. <https://doi.org/10.1002/mrm.27176>
- Fessler, J. A. (n.d.). *Michigan Image Reconstruction Toolbox*. Retrieved May 17, 2023, from <https://github.com/JeffFessler/mirt>
- Funai, A. K., Fessler, J. A., Yeo, D. T. B., Olafsson, V. T., & Noll, D. C. (2008). Regularized Field Map Estimation in MRI. *IEEE Transactions on Medical Imaging*, 27(10), 1484–1494. <https://doi.org/10.1109/TMI.2008.923956>
- Graedel, N. N., Hurley, S. A., Clare, S., Miller, K. L., Pruessmann, K. P., & Vannesjo, S. J. (2017). Comparison of gradient impulse response functions measured with a dynamic field camera and a phantom-based technique. 378.
- Graedel, N. N., Kasper, L., Engel, M., Nussbaum, J., Wilm, B. J., Pruessmann, K. P., & Vannesjo, S. J. (2021). Feasibility of spiral fMRI based on an LTI gradient model. *NeuroImage*, 245, 118674. <https://doi.org/10.1016/j.neuroimage.2021.118674>
- Hansen, M. S., & Sørensen, T. S. (2013). Gadgetron: An open source framework for medical image reconstruction: Gadgetron. *Magnetic Resonance in Medicine*, 69(6), 1768–1776. <https://doi.org/10.1002/mrm.24389>
- Inati, S. J., Naegele, J. D., Zwart, N. R., Roopchansingh, V., Lizak, M. J., Hansen, D. C., Liu, C., Atkinson, D., Kellman, P., Kozerke, S., Xue, H., Campbell-Washburn, A. E., Sørensen, T. S., & Hansen, M. S. (2017). ISMRM Raw data format: A proposed standard for MRI raw datasets. *Magnetic Resonance in Medicine*, 77(1), 411–421. <https://doi.org/10.1002/mrm.26089>
- Jaffray, A., Wu, Z. (Tim), Uludağ, K., & Kasper, L. (2022a). *Data Supplement: Open-source model-based reconstruction in Julia (ISMRM 2022)* [Data set]. Zenodo. <https://doi.org/10.5281/zenodo.6510021>

- Jaffray, A., Wu, Z., Uludağ, K., & Kasper, L. (2022b). Open-source model-based reconstruction in Julia: A pipeline for spiral diffusion imaging. *Proc. Intl. Soc. Mag. Reson. Med.* 30, 2435. <https://doi.org/10.58530/2022/2435>
- Kasper, L., Engel, M., Barmet, C., Haeberlin, M., Wilm, B. J., Dietrich, B. E., Schmid, T., Gross, S., Brunner, D. O., Stephan, K. E., & Pruessmann, K. P. (2018). Rapid anatomical brain imaging using spiral acquisition and an expanded signal model. *NeuroImage*, 168, 88–100. <https://doi.org/10.1016/j.neuroimage.2017.07.062>
- Kasper, L., Engel, M., Heinzle, J., Mueller-Schrader, M., Graedel, N. N., Reber, J., Schmid, T., Barmet, C., Wilm, B. J., Stephan, K. E., & Pruessmann, K. P. (2022). Advances in spiral fMRI: A high-resolution study with single-shot acquisition. *NeuroImage*, 246, 118738. <https://doi.org/10.1016/j.neuroimage.2021.118738>
- Kasper, L., Wu, Z., Jaffray, A., Kashyap, S., & Uludağ, K. (2023). Feasibility of spiral diffusion imaging on a clinical 3T MR system. *Proc. Intl. Soc. Mag. Reson. Med.* 31, 4164. <https://index.miramart.com/ISMRM2023/PDFfiles/4164.html>
- Knopp, T., Eggers, H., Dahnke, H., Prestin, J., & Senegas, J. (2009). Iterative Off-Resonance and Signal Decay Estimation and Correction for Multi-Echo MRI. *IEEE Transactions on Medical Imaging*, 28(3), 394–404. <https://doi.org/10.1109/TMI.2008.2006526>
- Knopp, T., & Grosser, M. (2021). MRIReco.jl: An MRI reconstruction framework written in Julia. *Magnetic Resonance in Medicine*, 86(3), 1633–1646. <https://doi.org/10.1002/mrm.28792>
- Lee, Y., Wilm, B. J., Brunner, D. O., Gross, S., Schmid, T., Nagy, Z., & Pruessmann, K. P. (2021). On the signal-to-noise ratio benefit of spiral acquisition in diffusion MRI. *Magnetic Resonance in Medicine*, 85(4), 1924–1937. <https://doi.org/10.1002/mrm.28554>
- Lin, C. Y., & Fessler, J. A. (2020). Efficient regularized field map estimation in 3D MRI. *IEEE Transactions on Computational Imaging*, 6, 1451–1458. <https://doi.org/10.1109/TCI.2020.3031082>
- NIFTI. (2003, September 2). *NIFTI Data Format* [National Institute of Mental Health Website]. Neuroimaging Informatics Technology Initiative. <https://nifti.nimh.nih.gov/>
- Pruessmann, K. P., Weiger, M., Börnert, P., & Boesiger, P. (2001). Advances in sensitivity encoding with arbitrary k -space trajectories: SENSE With Arbitrary k -Space Trajectories. *Magnetic Resonance in Medicine*, 46(4), 638–651. <https://doi.org/10.1002/mrm.1241>
- Robison, R. K., Li, Z., Wang, D., Ooi, M. B., & Pipe, J. G. (2019). Correction of B_0 eddy current effects in spiral MRI. *Magnetic Resonance in Medicine*, 81(4), 2501–2513. <https://doi.org/10.1002/mrm.27583>
- Sutton, B. P., Noll, D. C., & Fessler, J. A. (2003). Fast, iterative image reconstruction for MRI in the presence of field inhomogeneities. *IEEE Transactions on Medical Imaging*, 22(2), 178–188. <https://doi.org/10.1109/TMI.2002.808360>
- Uecker, M., Lai, P., Murphy, M. J., Virtue, P., Elad, M., Pauly, J. M., Vasanawala, S. S., & Lustig, M. (2014). ESPIRiT—an eigenvalue approach to autocalibrating parallel MRI: Where SENSE meets GRAPPA. *Magnetic Resonance in Medicine*, 71(3), 990–1001. <https://doi.org/10.1002/mrm.24751>
- Vannesjo, S. J., & Graedel, N. N. (2020). *MRI-gradient / GIRF*. MRI-gradient. <https://github.com/MRI-gradient/GIRF>
- Vannesjo, S. J., Graedel, N. N., Kasper, L., Gross, S., Busch, J., Haeberlin, M., Barmet, C., & Pruessmann, K. P. (2016). Image reconstruction using a gradient impulse response model for trajectory prediction: GIRF-Based Image Reconstruction. *Magnetic Resonance in Medicine*, 76(1), 45–58. <https://doi.org/10.1002/mrm.25841>

- Vannesjo, S. J., Haeberlin, M., Kasper, L., Pavan, M., Wilm, B. J., Barmet, C., & Pruessmann, K. P. (2013). Gradient system characterization by impulse response measurements with a dynamic field camera: Gradient System Characterization with a Dynamic Field Camera. *Magnetic Resonance in Medicine*, *69*(2), 583–593. <https://doi.org/10.1002/mrm.24263>
- Veldmann, M., Ehses, P., Chow, K., Nielsen, J., Zaitsev, M., & Stöcker, T. (2022). Open-source MR imaging and reconstruction workflow. *Magnetic Resonance in Medicine*, *88*(6), 2395–2407. <https://doi.org/10.1002/mrm.29384>
- Wilm, B. J., Barmet, C., Pavan, M., & Pruessmann, K. P. (2011). Higher order reconstruction for MRI in the presence of spatiotemporal field perturbations: Higher Order Reconstruction for MRI. *Magnetic Resonance in Medicine*, *65*(6), 1690–1701. <https://doi.org/10.1002/mrm.22767>
- Wilm, B. J., Nagy, Z., Barmet, C., Vannesjo, S. J., Kasper, L., Haeberlin, M., Gross, S., Dietrich, B. E., Brunner, D. O., Schmid, T., & Pruessmann, K. P. (2015). Diffusion MRI with concurrent magnetic field monitoring: Diffusion MRI with Concurrent Magnetic Field Monitoring. *Magnetic Resonance in Medicine*, *74*(4), 925–933. <https://doi.org/10.1002/mrm.25827>
- Wu, Z., Jaffray, A., Vannesjo, S. J., Uludağ, K., & Kasper, L. (2022). MR System Stability and Quality Control using Gradient Impulse Response Functions (GIRF). *Proc. Intl. Soc. Mag. Reson. Med.* *30*, 0641. <https://doi.org/10.58530/2022/0641>